

Statement for the CEMS results – January 2017

On the 12th January 2017 between 09:30-09:59 the CEMS instrument for the half hour average of Volatile Organic Carbon (VOC) recorded reading of 26.55 mg/m³. This time the Clinical Waste Incinerator CEMS FID equipment recorded unreasonable level of VOC, investigation carried out. A failed flexible sample hose that falsified the reading identified and repaired, consequently, this reading does not constitute a breach of any emission limit within the EPR Permit, nor does it indicate that the plant was performing abnormally.

A handwritten signature in blue ink, appearing to be 'GJH', is centered on a light blue rectangular background.

Signed:

Date: 16th January 2017